



Backflush Handle (Passive Aspiration)

This instrument incorporates a silicon backflush reservoir on the conventional Charles flute handle which helps in safe backflushing of the incarcerated tissue during passive aspiration of intraocular fluids.



Product Description

Product Code	Description	Qty. per box
MIPL/N7	Backflush Handle (Passive Aspiration)	1 pc.

Features

- Supplied unsterile
- User friendly design/ergonomic grip

Instructions for Use

The instructions for use and surgical technique are supplied in the form of a leaflet with instructions given in diagram form where appropriate . The international symbols used on the packaging are also explained.

Cleaning & Disinfection

For cleaning and disinfection of the device, kindly refer to the IFU supplied with the device or contact our customer care number



MADHU INSTRUMENTS PVT LTD

A-260, Okhla Industrial Area, Phase - 1, New Delhi - 110020, India

T: (+91-11) 4270 1030, 31 | E: sales@madhuinstruments.com | W: www.madhuinstruments.com

Customer Care No. +91 9810866176 