Maxiflo
Ophthalmic Cannulas

Anterior Chamber Maintainer

Anterior chamber maintainer is used for maintaining the anterior chamber during an ophthalmic surgical procedure.



Product Description

Product Code	Description	Qty. per box
MIPL/E4/20G	Anterior Chamber Maintainer (20G)	10 Pcs. (Sterile)
MIPL/E4/23G	Anterior Chamber Maintainer (23G)	10 Pcs. (Sterile)

Advantages

- The anterior chamber maintainer is connected to a separate saline bottle and the tip is inserted into the anterior chamber through the side port incision.
- In case there is any surge or the chamber becomes shallow, the saline from the bottle automatically fills the anterior chamber avoiding posterior capsule rupture or other surge related complications.
- Ultra thin wall needle for maximum flows.
- Disposable.

Maxiflo
Ophthalmic Cannulas

Instructions for Use

The instructions for use and surgical technique are supplied in the form of a leaflet with instructions given in diagram form where appropriate. The international symbols used on the packaging are also explained.

Sterilization

Products are sterilized through Ethylene Oxide (EO) according to a validated cycle.

madhu
INNOVATING POSSIBILITIES

MADHU INSTRUMENTS PVT LTD

A-260, Okhla Industrial Area, Phase - 1, New Delhi - 110020, India

T: (+91-11) 4270 1030, 31 | E: sales@madhuinstruments.com | www.madhuinstruments.com

Customer Care No. +91 9810866176 (for international queries), +91 7428186111 (for domestic queries)

Ref. No.: MIPL/SPB/E4/01, Rev. No. 05, Rev. Date: 29.09.2026